



Operation to ProPark-V3


檔案總管 → 我的電腦 (右鍵) → 電腦管理 → 裝置管理員 → 連接埠(查看)




RUN NovAtel CDU




Device → Open




EDIT or NEW




PORT (COM?) & 115200




OK -> OPEN




Logging Control (4th on the right)




Add log (far left)




Raw List
 RANGECMP
 15 sec




Add → close




Edit Recording Properties (far right)




Log file to be saved (such as desktop)




Start recording (if screen not shown, closing (x) the previous page)



move the page & check the epochs increased on RANGECOMPB



Stop recording



檢查 NB 桌面的檔案是否存在 且有儲存量

關閉 CDU 程式

關電源 拔各項接線